CellTec Activated Carbon Filters



CellTec activated carbon filters provide a high-value, low cost alternative to original manufacturer equipment filters used on many histology laboratory instruments.

The CellTec cartridges contain a high grade activated carbon which will absorb most organic and inorganic vapours including xylene, toluene, methanol, formalin, iso-propyl alcohol and industrial denatured alcohol.

To ensure air quality in the laboratory is maintained to safe and acceptable standards, it is essential that the carbon filters in equipment in contact with solvents and other potentially harmful products are replaced at the manufacturer's recommended interval.

KEY BENEFITS

Significant cost savings compared to instrument manufacturer equivalent

2 Low cost allows more frequent changing, and leads to improved staff safety

3 Identical size and fitting

Replacement Interval Guide and Part Number Reference

Manufacturer	Equipment	Replace Intervals*	CellPath Part No	Equivalent To
Leica	AutoStainer XL ST5010	3 months	XHA-0100-00A	Leica: 14047432273
	Multistainer ST5020	3 months	XHA-0100-00A	Leica: 14047432273
	Linear Stainer ST4040	3 months	XHA-0100-00A	Leica: 14047432273
	Coverslipper CV5000	3 months	XHA-0200-00A	Leica: 14046435914
	Tissue Processor TP1050	3 months	XHA-0300-00A	Leica: 14046135332
	Tissue Processor ASP300 / ASP300S	3 months	XHA-0400-00A	Leica: 14047634150
	Coverslipper CV5030 / Tissue Processor TP102 for organics	0 3 months	XHA-0500-00A	Leica: 14042230673
	Tissue Processor TP1020 for formaldehyde	3 months	XHA-0600-00A	Leica: 14042230674
	ASP6025 Tissue Processor	3 Months	XHA-0700-00A	Leica: 14049543860
Thermo Scientific	GLX Stainer	3 months	XHB-0200-00A	Shandon: 9990610
(Shandon)	HyperClean Fume Hood	3 months	XHB-0200-00A	Shandon: 9990610
	HyperClean (Enhanced)	4 weeks	XHB-0100-00A	Shandon: 9990612
	Consul / ClearVue Coverslipper	3 months	XHB-0200-00A	Shandon: 9990610
	Excelsior Tissue Processor	3 months	XHB-0200-00A	Shandon: 9990610
	Varistain Stainer / Gemini ES Stainer	3 months	XHB-0200-00A	Shandon: 9990610
	HyperClean, Excelsior	4 Weeks	XHB-0300-00A	Shandon: 9990612
Tissue-Tek	DRS Stainer	3 months	XHC-0100-00A	Sakura 2008
(Sakura)	SCA Coverslipper	3 months	XHC-0200-00A	Sakura 4775
	Prisma Range, VIP 6, Xpress x50	3 months	XHC-0300-00A	Sakura 6160
	Xpress x120	3 months	XHC-0400-00A	Sakura 7107
	VIP 5, VIP Jr, E-150, E-300	3 months	XHC-0500-00A	Sakura 4899

^{*}Replacement intervals are based on the current instrument equipment manufacturer guidelines. Replacement times will vary significantly depending on throughput.









